



T.C.  
BURDUR MEHMET AKİF ERSOY ÜNİVERSİTESİ  
AĞIZ VE DİŞ SAĞLIĞI HASTANESİ

MEDİKAL VAKUM VE SUCTION SİSTEMİ KONTROL FORMU

Doküman No

D.MC.FR.10

Yürürlük Tarihi

27.07.2023

Revizyon Tarih

00

Revizyon No

00

Sayfa No

1 / 1

CİHAZ ADI BLOK: GENEL	BİRİM	ÇALIŞMA ARALIĞI	KONTROL TARİH																																						
			01/...../20..	02/...../20..	03/...../20..	04/...../20..	05/...../20..	06/...../20..	07/...../20..	08/...../20..	09/...../20..	10/...../20..	11/...../20..	12/...../20..	13/...../20..	14/...../20..	15/...../20..	16/...../20..	17/...../20..	18/...../20..	19/...../20..	20/...../20..	21/...../20..	22/...../20..	23/...../20..	24/...../20..	25/...../20..	26/...../20..	27/...../20..	28/...../20..	29/...../20..	30/...../20..	31/...../20..								
VAKUM POMPALARI	*	E/H																																							
SUCTION POMPALARI	*	E/H																																							
VAKUM AKIŞI	*	E/H																																							
VAKUM BASINCI	BAR	450/600																																							
SUCTION AKIŞI	*	E/H																																							
SUCTION BASINCI	BAR	150/250																																							
GENEL TEMİZLİK KONTROLÜ	*	E/H																																							
KURUTUCU FİLTRE TEMİZLİĞİ	*	E/H																																							
VAKUM POMPALARI	*	E/H																																							
SUCTION POMPALARI	*	E/H																																							
KONTROL EDEN	KONTROL EDEN																																								

BIYOMEDİKAL TEKNİKER/TEKNİSYEN

HASTANE MÜDÜR/ MÜDÜR Y.